

IN LAB

LAB STATION 2

SCOPE: AGILENT 54622D
Ser. no. MV40003103

FUNC. GEN: AGILENT 33120A
Ser. no. 40010800

ITEM 2 Sine wave: 200mVpp , 2.5kHz , 0VDC offset
(Front panel readings)

Scope measurements:

Using cursors: 406mVpp
period = $400\mu\text{s} \Rightarrow f = 2.5\text{kHz}$


Auto. measurement: 405.9mVpp
 $f = 2.500\text{kHz}$

Discrepancy in amplitudes on func. generator meter and measured by scope.

Input Z of scope: $1\text{M}\Omega$.

Output Z of func. gen: 50Ω ALWAYS

Impedance mismatch causes the discrepancy in amplitudes.



The meter on the func. gen. reads correctly when it sees a 50Ω load. With a high Z load, set OUT TERM to HIGH Z to adjust meter reading.

With HIGH Z setting, func. gen. now reads
400mVpp

ITEM 3 Func. gen: $f_{max} = 15 \text{ MHz}$ $V_{max} = 10 \text{ Vpp}$ } into 50- Ω
 $f_{min} = 100 \mu\text{Hz}$ $V_{min} = 50 \text{ mVpp}$

Current scope config:

Triggering: edge, \uparrow , source CH1, DC coupled
 auto level mode

Input: DC coupled, Probe 1:1

Acquisitions: avg., 8 avgs.

ITEM 4 (a) Saved to INTERN-0: current sine wave + setup.

NOTE: Can recall trace only, setup only, or both.

Recalled trace + setup from INTERN-0.

(b) NOTE: Can delete + load files from floppy —
 UTILITY menu.

Saved sine wave + settings to QFILE-00 on
 floppy.

→ Saved as QFILE-00.SCP setup
 QFILE-00.TRC trace


Recalled from QFILE-00.

(c) Use QUICK PRINT to save screen as BITMAP
 to floppy, file PRINT-00.

NOTE: under UTILITY, PRINT CONFIG — setup
 to Print to Disk, FORMAT — BMP.

NOTE: Have screen display whatever you want to see on the printout.

Opened up PRINT_00.bmp on lab station computer using Paint, printed out put here.



ITEM 5

FFT display.

Scale = 20 dB/div Window = Hanning Offset = -55.0 dBV
Span = 20 kHz Center = 10 kHz Sample rate = 200 kSa/s

Peak line at: $f = 2.50 \text{ kHz}$ Ampl = -16.88 dBV

Very small peak at 5.0 kHz , -83.13 dBV

The func. gen. does not produce a pure sine wave — other freq. components are present. But, they are mostly in the noise floor — the line at 5 kHz is 66.25 dB below the fundamental.

Comparison with theory: (per lab).

Line at $f_0 = 2.5 \text{ kHz}$ should have amplitude

$$A = \frac{400 \text{ mV}_{pp}}{2} = 200 \text{ mV. (The FFT display}$$

is one-sided : amplitude of fundamental is

A , not $A/2$.)

~~$$A_{dBV} = 20 \log \frac{200 \text{ mV}}{1 \text{ V}_{rms}}$$~~

Units of measurement on scope are dBV, which

is:

$$\text{dBV} = 20 \log \left(\frac{A}{1 \text{ V}_{rms}} \right)$$


NOTE: Ref. is 1 V_{rms}

$$A = 200 \text{ mV} \Rightarrow A = \frac{200 \text{ mV}}{\sqrt{2}} = 141.42 \text{ mV}_{rms}$$

$$\therefore A_{dBV} = 20 \log (141.42 \times 10^{-3}) = \underline{\underline{-16.99 \text{ dBV}}}$$

Measured value: -16.88 dBV.



ITEM 6 PRINT_01.bmp



The lines have non-zero width because of the window — we are taking the spectrum of $(\cos 2\pi f_0 t)(u(t) - u(t-T))$, $T = \text{window width}$. $|\mathcal{F}[u(t) - u(t-T)]| = T \text{sinc}(Tf)$.

ITEM 7

Square wave: PRINT_02 FFT: PRINT_03



FFT settings for square wave:

Scale = 20 dB/div Offset = -63.4 dB Window = Hanning
 Span = 100 kHz Center = 50 kHz Rate = 1.00 MSa/s


Note that the spectrum should have lines only at odd harmonics (10 kHz, 30 kHz, etc.). There are small lines at even harmonics because it is not a perfect square wave.

	Line freq. (kHz)	Meas. Ampl. (dBV)	Meas. Ampl. (V _{op})	Theor. Ampl. (V _{op})	$(2 \cdot \frac{2A}{\pi n})$
n=1	10	-23.75	91.84 mV	127.32 mV	$(\frac{400m}{\pi})$
	20	-86.25	68.87 μV	0	
n=3	30	-32.50	33.53 mV	42.44 mV	$(\frac{400m}{3\pi})$
	40	-86.25	68.87 μV	0	
n=5	50 (49.9)	-37.50	18.85 mV	25.46 mV	$(\frac{400m}{5\pi})$
	60	-86.25	68.87 μV	0	
n=7	70 (69.9)	-40.00	14.14 mV	18.19 mV	$(\frac{400m}{7\pi})$
	80	-86.25	68.87 μV	0	
n=9	90	-42.50	10.61 mV	14.15 mV	$(\frac{400m}{9\pi})$

Conversion: $\text{dBV} = 20 \log\left(\frac{A_{\text{rms}}}{1V_{\text{rms}}}\right) \Rightarrow A_{\text{rms}} = 10^{\frac{\text{dB}}{20}}$



$\Rightarrow A = \sqrt{2} \cdot 10^{\frac{\text{dB}}{20}}$

ITEM 8 RC filter



V_{in} : square, $f = 10$ kHz, $A = 200mV_{pp}$, 0V offset


1st plat: sq. wave out of filter
2nd plat: measurement of time const.



ITEM 9 Time constant measurement — use delayed sweep.

Total swing: $\Delta = 200 \text{ mV}$

Use Y cursors to find 10% and 90% levels, then X cursors to measure rise time:



$$t_r = \frac{23.6}{\cancel{2.4}} \mu\text{s}$$


$$\left. \begin{aligned} 0.1\Delta &= \Delta(1 - e^{-t_1/\tau}) \\ 0.9\Delta &= \Delta(1 - e^{-t_2/\tau}) \end{aligned} \right\} t_2 - t_1 = \tau \ln(9)$$

$$\Rightarrow t_r = \tau \ln(9)$$

~~$$\tau = \frac{t_r}{\ln 9} = \frac{22.4 \mu\text{s}}{\ln 9} = 10.2 \mu\text{s}$$~~

$$\tau = \frac{t_r}{\ln(9)} = \frac{23.6 \mu\text{s}}{\ln 9} = 10.74 \mu\text{s} \quad \text{Expected } \tau.$$

FFT of output: Scale: 20dB/div Offset: -60dB Window Hanning
Span: 100kHz Center: 50kHz Rate: 1Ms/s



	Line freq (kHz)	Meas Ampl (dBV)	Meas. Ampl (V _{o-p})	Theoretical Ampl. (V _{o-p})
n=1	10	-22.50	106.05 mV	107.81 mV
2	20	-91.25	39 μV	0
3	30	-37.50	18.86 mV	19.89 mV
4	40	-87.50	60 μV	0
5	50	-45.62	7.41 mV	7.72 mV
6	60	-87.50	60 μV	0
7	70	-50.62	4.16 mV	4.03 mV
8	80	-85.63	74 μV	0
9	90	-55.00	2.52 mV	2.46 mV

$$A = \sqrt{2} 10^{\text{dBV}/20}$$

p. 23